



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
-----------------	-------------	----------------------	---------------------	------------------

10/637,221

08/08/2003

Randall M. Smith

400290

3334

27717 7590 06/24/2009
SEYFARTH SHAW LLP
131 S. DEARBORN ST., SUITE 2400
CHICAGO, IL 60603-5803

EXAMINER

LAURITZEN, AMANDA L

ART UNIT

PAPER NUMBER

3737

MAIL DATE

DELIVERY MODE

06/24/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary	Application No. 10/637,221	Applicant(s) SMITH ET AL.	
	Examiner Amanda L. Lauritzen	Art Unit 3737	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 12 June 2008.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-46 and 48-53 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-46 and 48-53 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date. _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

Art Unit: 3737

DETAILED ACTION

This action is in response to communications filed 12 June 2008. Amendments to the claim(s) are not interpreted to introduce new matter.

Response to Arguments

Applicant's arguments have been fully considered and are persuasive; therefore the rejection under 35 U.S.C. 103(a) over Carr in view of Meaney '943 is withdrawn. However, upon further consideration, new grounds of rejection is made in view of Carr (US 5,983,124).

Priority

1. Applicant's claim for the benefit of a prior-filed application under 35 U.S.C. 119(e) or under 35 U.S.C. 120, 121, or 365(c) is acknowledged.

Claim Objections

2. Claim 1 is objected to as it should read "a support system *adapted* to support..." at line 4. Additionally, claim 46 omits the term adapted in the claim. Appropriate correction is required.

Drawings

The drawings are objected to under 37 CFR 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the top planar envelope of at least claim 41 must be shown or the feature(s) canceled from the claim(s). Additionally, the specification describes a scan envelope, 39, at paragraph [0048]-[0049] and [0066], but this

Art Unit: 3737

feature is not depicted in the drawings. No new matter should be entered in the corrected drawing sheets.

Corrected drawing sheets in compliance with 37 CFR 1.121(d) are required in reply to the Office action to avoid abandonment of the application. Any amended replacement drawing sheet should include all of the figures appearing on the immediate prior version of the sheet, even if only one figure is being amended. The figure or figure number of an amended drawing should not be labeled as “amended.” If a drawing figure is to be canceled, the appropriate figure must be removed from the replacement sheet, and where necessary, the remaining figures must be renumbered and appropriate changes made to the brief description of the several views of the drawings for consistency. Additional replacement sheets may be necessary to show the renumbering of the remaining figures. Each drawing sheet submitted after the filing date of an application must be labeled in the top margin as either “Replacement Sheet” or “New Sheet” pursuant to 37 CFR 1.121(d). If the changes are not accepted by the examiner, the applicant will be notified and informed of any required corrective action in the next Office action. The objection to the drawings will not be held in abeyance.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Art Unit: 3737

3. Claims 1, 2, 6, 7, 9, 10, 17-20, 21, 24-38, 41-46 and 49-53 are rejected under 35 U.S.C. 103(a) as being unpatentable over Carr (US 5,983,124)

Carr discloses an apparatus and associated method for screening or diagnosing cancer in the breast of a patient, comprising: a support system for supporting the patient's breast in a fixed position; a microwave assembly including an antenna, source, receiver, and inherent processor; and an orientation system for orienting the surface of the breast in known positions with respect to the anatomy of the patient and locations of the antenna (microwave-transparent padded scan plate shown in Figs. 1 and 6 enables orientation of the breast in known positions relative to the position indicia and with respect to both the patient anatomy and antenna(s) located within the probe as in col. 4, lines 20-35; col. 5, lines 10-35; col. 9, lines 43-60; also col. 2, line 61 – col. 3, line 10, in which positioning is with respect to antenna(s)).

The method and system of Carr substantially includes all features of the claimed invention, but is not specific to providing a patient in a prone position on a table; however it would have been obvious to one ordinarily skilled in the art at the time of invention to modify the supine positioning of the patient, as disclosed in the system of Carr with a patient positioned in prone, such that the breast is positioned above the microwave-transparent padded scan plate of Fig. 1, rather than below, as it is known to skilled artisans to position patients in any of a number of positions for breast imaging scans, including both prone and supine positioning. Modifying the invention of Carr for prone patient positioning is a simple rearrangement of parts, as the method steps and technologies employed are applicable regardless of the orientation of the patient (and/or design of the patient supporting structure in how it accommodates or receives the patient). The system and method in the disclosure of Carr are identical to the claimed invention,

Art Unit: 3737

save for the basic modification required to image in a prone rather than a supine position, both of which have their advantages that are well-established in medical imaging of the breast tissue; for example, it is understood that the relative comfort of the patient is enhanced with apparatuses that accommodate positioning in the prone position.

Regarding claim 7, Carr discloses a microwave-transparent scan plate (shown in Figs. 1 and 6) but does not disclose the dielectric constant of the plate to be within the range of 1.7-9; however other materials of the apparatus are disclosed within that range (e.g., the probe at col. 8, lines 46-54). Because the plate is microwave-transparent and appropriate for imaging, it is understood that this component of the apparatus is also within this dielectric range.

Regarding claim 10, Examiner understands an air gap of less than 3 mm to be provided between the antenna and scan plate of the apparatus of Carr as the antenna is disclosed to make intimate contact (col. 5, lines 32-35), which is most broadly interpreted to encompass close proximity and/or material touching.

Regarding claim 21, Carr does not expressly disclose digital camera viewing means, but it would have been obvious to provide a digital camera imaging system, as digital imaging system are well known in the art to provide convenience and flexibility for image acquisition, processing, and viewing.

Regarding claim 25, the support member is taken to be the grid/scan plate (of Figs. 1 and 6) of the Carr reference.

Regarding claims 32-33, archiving displayed image and patient data are take as image and data storage means as disclosed in Carr.

Art Unit: 3737

4. Claims 3-5, 8, 21, 22 are rejected under 35 U.S.C. 103(a) as being unpatentable over Carr, as applied to claim 2 above, further in view of Meaney et al. (US 2004/0077943).

Carr discloses all limitations of the invention as substantially claimed including examination of the armpit area (auxillary gland at col. 2, line 65) and as detailed above, but does not disclose incorporating an optical camera in the microwave imaging system; however, in the same field of microwave array imaging, Meaney et al. teach acquiring optical images that are to be overlaid with the microwave scan images (par. 14, in which microwave images are “spatially co-registered” with a 3-D optical image; see also claim 70 for overlaying with a 3-D optical image). It would have been obvious to one of ordinary skill in the art at the time of invention to combine the optical imaging system with the microwave imaging system as taught by Meaney for the purpose of corresponding microwave-detected internal abnormalities with a specific visual reference of the breast exterior (for motivation, see par. 54 of Meaney et al).

Regarding claim 21, neither Carr nor Meaney expressly disclose digital camera viewing means, but it would have been obvious to provide a digital camera for the optical imaging means disclosed with the microwave/optical imaging system of Meaney as digital imaging systems are well known in the art to provide convenience and flexibility for image acquisition, processing, and viewing.

Regarding claim 22, the antenna contained within the probe of Carr is disclosed to move along coordinates (refer to Fig. 6 for coordinate grid; see also col. 9, lines 43-60), but this movement is not disclosed as provided by a motorized system; however, movement of the antenna disclosed by Meaney is provided with an actuator/drive shaft (paras. 12-13). Since movement of the antenna of Meaney does not require physical placement on the part of the

Art Unit: 3737

medical examiner, the drive shaft inherently includes some sort of motorized system (additionally, a motor is a simple means of actuation that is well known in the art). It would have been obvious to include a motorized actuator as described by Meaney for movement of the antenna along coordinates as disclosed by Carr in order to enhance accuracy and precision by automating placement of the probe.

5. Claims 11, 12, 13-16, and 39-40 are rejected under 35 U.S.C. 103(a) as being unpatentable over Carr '124 in view of Haddad et al. (US 6,454,711).

Carr discloses all features of the invention as substantially claimed but does not include microwave absorbing material, but in the same field of endeavor Haddad et al. disclose microwave absorbing material (col. 3, line 27). It would have been obvious to incorporate use of a microwave absorbing material for the purpose of reducing residual crosstalk between the antennas of the apparatus of Carr (for motivation, see Haddad col. 3, lines 27-28).

6. Claim 23 and 48 are rejected under 35 U.S.C. 103(a) as being unpatentable over Carr '124 in view of Horton et al. (US 5,168,514).

Carr discloses all features of the invention as substantially claimed but does not include an adjustable upper surface section of the examination table that enables the patient to sit in an upright position; however, Horton et al. disclose adjustable back-rests and other surfaces as part of an examination table that provide a patient with support in a seated position while undergoing medical procedures in which breast tissue is targeted (col. 2, lines 44-47). It would have been obvious to one of ordinary skill in the art to provide the breast examination table of Carr with

Art Unit: 3737

adjustable support members disclosed by Horton in order to position the patient such that the targeted anatomy is fully accessible to the medical examiner.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Amanda L. Lauritzen whose telephone number is (571)272-4303. The examiner can normally be reached on Monday - Friday, 8:30am - 5:00pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Brian L. Casler can be reached on (571) 272-4956. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Amanda L. Lauritzen/
Examiner, Art Unit 3737

/BRIAN CASLER/
Supervisory Patent Examiner, Art Unit
3737